



```
=====
Acq. Operator   : SYSTEM                               Seq. Line :   12
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                               Location  : P1-C-11
Injection Date  : 10/5/2021 1:25:22 PM                Inj       :    1
                                                    Inj Volume: 1.000 µl
```

```
Acq. Method     : C:\Users\Public\Documents\ChemStation\1\Data\10052101\010042101 2021-10-05 09-56-30\Purity_195-315.M
Last changed    : 4/29/2021 10:41:07 AM by SYSTEM
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
Last changed    : 9/30/2021 9:18:37 AM by SYSTEM
                (modified after loading)
```

```
Method Info     : OPERATOR: Organic QC lab
                General gradient method starting @100% water for 1 min @0.3 ml/min,the 100% Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system , Agilent 6125 quadrupole MS detector.
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

