

Acq. Operator : SYSTEM

Seq. Line :

**AccuStandard**
Leader in Analytical Reference Standards

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2

Location : P2-A-10

Injection Date : 10/18/2022 4:16:33 PM

Inj : 1

Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\10182201\10182201 2022-10-18 14-03-20
\Purity_195-315.M

Last changed : 3/23/2022 11:04:54 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 3/23/2022 11:04:54 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Phenylurea Primary Dilution @ Varied Conc. in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

