Print of window 38: Current Chromatogram(s)
Data File : C:\USERS\P...TATION\1\DATA\020924\0209224 2024-02-09 07-22-31\18- KD.D
Sample Name : M-537

==========

Acq. Operator : SYSTEM Seq. Line : 18
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P2-A-01
Injection Date : 2/9/2024 12:44:13 PM Inj : 2
Inj Volume : 2.000 µl
Different Inj Volume from Sample Entry! Actual Inj Volume : 0.500 µl
Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\020924\0209224 2024-02-09 07-22-31 \PFAA.M
Last changed : 1/26/2021 10:30:23 AM by SYSTEM
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M
Last changed : 2/6/2023 2:02:18 PM by SYSTEM
Method Info : OPERATOR: Organic QC lab
General gradient method starting @100% water for 1 min @0.3 ml/min,the 100% Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system , Agilent 6125 quadrupole MS detector.

Current Chromatogram(s)