



```
=====
Acq. Operator   : SYSTEM                               Seq. Line :    2
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                               Location  : P1-E-01
Injection Date  : 6/29/2022 6:12:04 AM                Inj       :    1
                                                    Inj Volume: 2.000 µl
```

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.500 µl

```
Acq. Method     : C:\Users\Public\Documents\ChemStation\1\Data\06292201\06292201 2022-06-29
                  05-54-01\Diquat-Paraquat-Chlormequat.M
```

Last changed : 5/28/2020 9:44:05 AM by SYSTEM

```
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
```

Last changed : 3/23/2022 11:04:54 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

