

File :D:\MassHunter\GCMS\1\data\20221215\04- (JP).D  
Operator : Organic QC Lab  
Acquired : 15 Dec 2022 15:31 using AcqMethod Hy\_M551.1-MLPC\_2ulinj.M  
Instrument : GCMS 8  
Sample Name: M-551.1-LPC  
Misc Info in mtbe  
Vial Number: 5



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

