

File :Q:\GCMS-08 Minimal\DATA\20191017\07 - .D
Operator : KM / JP / ZA
Acquired : 17 Oct 19 02:56 pm using AcqMethod KERLA-VOC-SV.M
Instrument : GCMS 8
Sample Name: M-552.2A-08
Misc Info : Monochloroacetic acid in MtBE
Vial Number: 20



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

