



=====  
 Acq. Operator : SYSTEM Seq. Line : 22  
 Sample Operator : SYSTEM  
 Acq. Instrument : LCMS-2 Location : P1-C-11  
 Injection Date : 5/11/2021 3:17:41 PM Inj : 1  
 Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\05112101\05112101 2021-05-11 08-39-51  
 \Purity\_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity\_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%  
 Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse  
 plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system  
 , Agilent 6125 quadrupole MS detector.

Sample Info : DILUTED TO 100UG/ML

Additional Info : Peak(s) manually integrated

### Current Chromatogram(s)

