



=====
 Acq. Operator : SYSTEM Seq. Line : 13
 Sample Operator : SYSTEM
 Acq. Instrument : LCMS-2 Location : P2-D-10
 Injection Date : 1/7/2022 2:54:42 PM Inj : 1
 Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\01072201\01062201 2022-01-07
 11-07-21\Purity_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 4/29/2021 10:41:07 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
 Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax
 Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity
 II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Chlorinated Acids by HPLC @ 1000ug/mL in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

