

File :Z:\QC Data\GCMS-07 Minimal\DATA\01162001\22- .D  
Operator : Organic QC department  
Acquired : 16 Jan 20 03:49 pm using AcqMethod KERLA-2017.M  
Instrument : 5977B MSD  
Sample Name: M-608-1  
Misc Info : in MeOH  
Vial Number: 114



Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

