

File :Q:\GCMS-08 Minimal\DATA\20190530\38- .D
Operator : JP/KM
Acquired : 01 Jun 2019 01:35 using AcqMethod KERLA-2Split-2uL.M
Instrument : GCMS 8
Sample Name: M-612
Misc Info : RT: Chlorinated Hydrocarbons in Isooctane
Vial Number: 128



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

