


File :D:\MassHunter\GCMS\1\DATA\05182001\02- -KD.D  
Operator : Organic QC department  
Acquired : 18 May 2020 06:15 using AcqMethod KERLA-2017\_Short\_15Min.M  
Instrument : 5977B MSD  
Sample Name:M-622-12  
Misc Info :  
Vial Number: 114

 **AccuStandard**<sup>®</sup>  
Leader in Analytical Reference Standards  
Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

