

File :Q:\GCMS-08\DATA\20190222\04- D
Operator : JP/KM
Acquired : 22 Feb 19 10:55 am using AcqMethod KERLA-2017_S50.M
Instrument : GCMS 8
Sample Name: M-622-23
Misc Info : Stirophos in Hexane
Vial Number: 41



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

