

File :M:\GCMS-07 Minimal\DATA\04272201\13- ALR.D  
Operator : Organic QC department  
Acquired : 27 Apr 2022 15:24 using AcqMethod KERLA-VOC-SV\_S100.M  
Instrument : 5977B MSD  
Sample Name: M-625-MOD  
Misc Info : Method 625 Modificationn DCM  
Vial Number: 134



**AccuStandard®**

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

