

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-08 Minimal\DATA\20230220
... \34- =JP=.D
Operator : Organic QC Lab
Instrument : GCMS 8
Acquired : 21 Feb 2023 00:28 using AcqMethod Hy_PAH_15cm_S5.M
Sample Name: M-625A
Misc Info : in meoh



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

