

File :Q:\GCMS-07 Minimal\DATA\05192001\13- KM.D
Operator : Organic QC department
Acquired : 19 May 2020 20:59 using AcqMethod KERLA-2017.M
Instrument : 5977B MSD
Sample Name: M-625C-2-10X
Misc Info : Pentachlorophenol in MeCl2
Vial Number: 140



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

