



Acq. Operator : SYSTEM

Seq. Line : 2

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2

Location : P1-F-07

Injection Date : 5/31/2022 3:14:23 PM

Inj : 2

Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\05312201\05312201 2022-05-31
14-54-27\Purity_195-315.M

Last changed : 3/23/2022 11:04:54 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 5/16/2022 12:35:31 PM by SYSTEM
(modified after loading)

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax
Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity
II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Methiocarb in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

