

File : Q:\GCMS-08 Minimal\DATA\20220712\33 -jopo.D
Operator : Organic QC Lab
Acquired : 12 Jul 2022 19:40 using AcqMethod Hy_VOC-SV_24cm_S100.M
Instrument : GCMS 8
Sample Name: M-8020-SS
Misc Info in meoh
Vial Number: 133



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

