

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-04 Minimal\DATA\112122\0
... 7-KM.D
Operator :
Instrument : GCMS4
Acquired : 21 Nov 2022 19:52 using AcqMethod VOC-100 SPLIT #H2 20MIN175DEG.M
Sample Name: M-8021B-SS-100X
Misc Info : Surrogate Standards Mix in MeOH



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

