

File :D:\MassHunter\GCMS\1\data\09162022\21 - SAK.D

Operator :

Acquired : 16 Sep 2022 16:06 using AcqMethod Hy SV DB-5MS UI20M 0.18 40 split Time 14 MIN .M

Instrument : GCMS 9

Sample Name: M-8041-IS-10X

Misc Info : IS mix in Isopropanol

Vial Number: 121

Column: DB-5MS UI, 20m x 0.18mm ID x 0.18 $\mu$ m

Oven Program: 45°C, 13°C/min to 290°C (hold 5min)

GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1 $\mu$ l inj.; GC/MS; Injector temp 270°C

