

File :D:\MassHunter\GCMS\1\data\06141802\18-  
Operator K.D  
Acquired : 14 Jun 2018 22:30 using AcqMethod 2017\_Short\_15Min.M  
Instrument : 5977B MSD  
Sample Name: M-8081-IS  
Misc Info Quintozene in Acetone  
Vial Number: 65

Column: DB-5, 30m, 0.25 ID, 0.25um  
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.  
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow  
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

