

File : L:\QC Data\GCMS-06 Minimal\DATA\050823\15 - SAK.D  
Operator : Organic QC Lab  
Acquired : 08 May 2023 12:54 using AcqMethod CHICK\_2019\_13MIN\_2split\_2uL.M  
Instrument : GCMS 6  
Sample Name : M-8082-SSA-WL-10ML  
Misc Info :  
Vial Number : 109



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

