

File :Q:\GCMS-08 Minimal\DATA\20230227\34- -JJP-.D
Operator : Organic QC Lab
Acquired : 28 Feb 23 05:04 pm using AcqMethod Hy_PAH_24cm_S100.M
Instrument : GCMS 8
Sample Name: M-8120-03
Misc Info : 1,3-Dichlorobenzen; in hexane
Vial Number: 14



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; injector 270°C

