

File :Q:\GCMS-06 Minimal\DATA\022421\34- -DC.D
Operator : Organic QC Lab
Acquired : 25 Feb 21 01:38 am using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: M-8141A-SS
Misc Info : M-8141A-SS in Acetone
Vial Number: 51



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

