

```

=====
Acq. Operator   : SYSTEM                      Seq. Line :   20
Acq. Instrument : LCMS-1                      Location  : P1-C1
Injection Date  : 9/2/2021 5:36:09 PM        Inj       :    1
                                                Inj Volume: 1.000 µl
    
```



```

Acq. Method     : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 5/21/2021 4:03:06 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 8/16/2021 2:49:45 PM by SYSTEM
                 (modified after loading)
Method Info     : Operator: Organic QC lab
                 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                 Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                 :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                 particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                 MS Detector. No UV Reference.
    
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

