File :Q:\GCMS-04\DATA\111418\05-
Operator : RFXM
Acquired : 14 Nov 2018 9:46 using AcqMethod VOC-100SPLIT DECKER.M
Instrument : GCMS4
Sample Name: M-8240A
Misc Info :
Vial Number: 30

AccuStandard®

Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C