

File : Q:\GCMS-04 Minimal\DATA\110221B\21- YX.D  
Operator : Organic QC Lab  
Acquired : 3 Nov 2021 1:25 using AcqMethod VOC-200SPLIT DECKER.M  
Instrument : GCMS4  
Sample Name: M-8260A-B-SS  
Misc Info : Surrogate Standard Mix in MeOH  
Vial Number: 86



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm  
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)  
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

