

File :B:\QC Data\GCMS-08 Minimal\DATA\20190815\20-.D  
Operator : JJP / ZA  
Acquired : 15 Aug 2019 16:44 using AcqMethod KERLA-2017.M  
Instrument : GCMS 8  
Sample Name: M-8270-07-ASL  
Misc Info : Pyridine mix in Acetone  
Vial Number: 11



Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

