

=====

Acq. Operator : SYSTEM	Seq. Line : 1
Sample Operator : SYSTEM	
Acq. Instrument : LCMS-2	Location : P2-D-02
Injection Date : 3/19/2020 3:50:59 PM	Inj : 1
	Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.100 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\03192001\02192001 2020-03-19 15-50-04
\Purity_195-315.M

Last changed : 12/3/2019 1:25:24 PM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 3/17/2020 3:54:55 PM by SYSTEM

(modified after loading)

Method Info : General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

