

=====
Acq. Operator : SYSTEM Seq. Line : 2
Acq. Instrument : LCMS-1 Location : P2-D3
Injection Date : 3/27/2019 6:48:57 AM Inj : 2
Inj Volume : 1.000 µl
Acq. Method : D:\DATA\03271601\03271601 2019-03-27 06-31-15\DNPH.M
Last changed : 3/27/2019 6:31:15 AM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\CarbamatesPT.M
Last changed : 3/16/2017 10:42:31 AM by SYSTEM
Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV
Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

