

=====
Acq. Operator : SYSTEM Seq. Line : 8
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P2-E-04
Injection Date : 6/13/2022 11:01:46 AM Inj : 1
Inj Volume : 2.000 µl



Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\06132201\06132201 2022-06-13 08-48-48
\PFAA.M

Last changed : 1/26/2021 10:30:23 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 5/16/2022 12:35:31 PM by SYSTEM
(modified after loading)

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

