



=====

Acq. Operator : SYSTEM	Seq. Line : 4
Sample Operator : SYSTEM	
Acq. Instrument : LCMS-2	Location : P2-D-01
Injection Date : 4/8/2021 11:01:11 AM	Inj : 1
	Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 2.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\04082101\04072101 2021-04-08 10-03-30
\Purity_195-315.M

Last changed : 5/8/2020 10:12:19 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M

Last changed : 3/29/2021 1:57:49 PM by SYSTEM
(modified after loading)

Method Info : OPERATOR: Organic QC lab
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse
plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system
, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

