

=====

Acq. Operator : SYSTEM	Seq. Line : 4
Acq. Instrument : LCMS-1	Location : P1-C6
Injection Date : 5/24/2021 3:10:42 PM	Inj : 1
	Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 5/21/2021 4:03:06 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 5/25/2021 9:11:15 AM by SYSTEM
(modified after loading)

Method Info : Operator: Oraganic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV
Reference.



Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

