Print of window 38: Current Chromatogram(s)
Data File : D:\DATA\12181901\32- .D
Sample Name : M-8330-06-0.1X

==========

Acq. Operator : SYSTEM Seq. Line : 31
Acq. Instrument : LCMS-1 Location : P2-A3
Injection Date : 12/18/2019 7:34:28 PM Inj : 1
Inj Volume : 1.000 µl
Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 10/18/2019 4:18:14 PM by SYSTEM
Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Sample Info : Nitrobenzene in AcCN/MeOH (50/50)

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)