

=====  
Acq. Operator : SYSTEM  
Acq. Instrument : LCMS-1  
Injection Date : 5/7/2021 2:16:50 PM  
Seq. Line : 8  
Location : P1-A8  
Inj : 1  
Inj Volume : 1.000 µl



Acq. Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 5/8/2020 11:20:44 AM by SYSTEM  
Analysis Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 5/7/2021 2:11:41 PM by SYSTEM  
(modified after loading)  
Method Info : Operator: Oraganic QC lab  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

