

```

=====
Acq. Operator   : SYSTEM                Seq. Line :    3
Acq. Instrument : LCMS-1                Location  : P1-F3
Injection Date  : 5/26/2021 3:54:00 PM Inj       :    1
                                       Inj Volume : 1.000 µl

Acq. Method     : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 5/21/2021 4:03:06 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 5/27/2021 11:27:39 AM by SYSTEM
                 (modified after loading)

Method Info     : Operator: Oraganic QC lab
                 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                 Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                 /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
                 Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV
                 Reference.

```



Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

