

=====

Acq. Operator	: SYSTEM	Seq. Line	: 5
Acq. Instrument	: LCMS-1	Location	: P1-F6
Injection Date	: 2/3/2022 10:02:45 AM	Inj	: 1
		Inj Volume	: 1.000 µl



Acq. Method	: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed	: 11/5/2021 2:37:17 PM by SYSTEM
Analysis Method	: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed	: 1/27/2022 4:16:54 PM by SYSTEM
	(modified after loading)
Method Info	: Operator: Organic QC lab
	Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadruple MS Detector. No UV Reference.

Sample Info : Erythritol tetranitrate in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

