


File : B:\QC Data\GCMS-08 Minimal\DATA\20191007\Z01-D  
Operator : ZA / KM / JP  
Acquired : 08 Oct 2019 12:01 using AcqMethod KERLA-2017.M  
Instrument : GCMS 8  
Sample Name : M-8330-ADD-50  
Misc Info : ETHYL CENTRALITE IN ACCN/MEOH  
Vial Number : 131

 **AccuStandard®**  
Leader in Analytical Reference Standards  
Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

