Print of window 38: Current Chromatogram(s)
Data File : D:\DATA\12142021\10- KM.D
Sample Name : M-8330-ADD-6

==========

Acq. Operator : SYSTEM
Seq. Line : 4
Acq. Instrument : LCMS-1
Location : P1-A7
Injection Date : 12/14/2021 3:31:13 PM
Inj : 1
Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 11/5/2021 2:37:17 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 12/14/2021 9:28:32 AM by SYSTEM
(modified after loading)

Method Info : Operator: Organic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Sample Info : Nitroguanidine @ 100 ug/mL in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)