

File :D:\MassHunter\GCMS\1\DATA\092721\03- ALR.D  
 Operator : Organic QC Lab  
 Acquired : 27 Sep 2021 10:18 using AcqMethod CHICK\_2019\_S100.M  
 Instrument : GCMS 6  
 Sample Name: M-8410-IS  
 Misc Info : 1-Fluoronaphthalene & p-Terphenyl-d14 in DCM  
 Vial Number: 145

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
 Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
 GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
 Split 20:1, 1µl inj.; GC/MS; Injector 270°C

