


File : Q:\GCMS-04\DATA\030419\05-.D
Operator : RFXM
Acquired : 4 Mar 2019 17:57 using AcqMethod VOC-SCAN 24 DECKER.M
Instrument : GCMS4
Sample Name: M-GRO-CAL-IS-01-R1
Misc Info :
Vial Number: 38



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C



Not in mix - Diisopropyl ether (108-20-3), EtBE (637-92-3), TAME (994-05-8), Methanol (67-56-1)