

File :L:\QC Data\GCMS-08 Minimal\DATA\20220908\21 - SAK.D
Operator : Organic QC Lab
Acquired : 08 Sep 2022 17:29 using AcqMethod Hy_Pest_15cm_S2_3ulinj.M
Instrument : GCMS 8
Sample Name: M-PHE
Misc Info : Phthalate esters mix in Acetone
Vial Number: 131



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

