


File :L:\QC Data\GCMS-06 Minimal\DATA\111419B\22-.D  
Operator : ZA / KM / YX  
Acquired : 14 Nov 2019 20:42 using AcqMethod CHICK\_2019\_VOC-SV\_S100.M  
Instrument : GCMS 6  
Sample Name: P-010S-10X  
Misc Info :  
Vial Number: 134

 **AccuStandard®**  
Leader in Analytical Reference Standards  
Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

