



=====

Acq. Operator	: SYSTEM	Seq. Line	: 2
Acq. Instrument	: LCMS-1	Location	: P2-C3
Injection Date	: 6/19/2018 11:01:58 AM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : D:\DATA\06181801\060181801 2018-06-19 10-42-29\Purity195_315.M
Last changed : 6/19/2018 10:42:29 AM by SYSTEM
Analysis Method : C:\Chem32\1\Methods\Purity195_315.M
Last changed : 6/18/2018 9:35:46 AM by SYSTEM
(modified after loading)

Method Info : Operator: GA
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

