

File : Q:\GCMS-07 Minimal\DATA\07082001\92- YX.D
Operator : Organic QC department
Acquired : 10 Jul 2020 11:06 using AcqMethod KERLA-2Split-2uL.M
Instrument : 5977B MSD
Sample Name: P-024S-2
Misc Info : o,p'-DDD in Isooctane
Vial Number: 112



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

