



=====  
Acq. Operator : SYSTEM Seq. Line : 5  
Sample Operator : SYSTEM  
Acq. Instrument : LCMS-2 Location : P1-E-01  
Injection Date : 6/24/2020 5:59:43 PM Inj : 1  
Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\06242001\06242001 2020-06-24 16-43-02  
\Purity\_195-315.M

Last changed : 5/8/2020 10:12:19 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 6/24/2020 12:11:23 PM by SYSTEM  
(modified after loading)

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100%  
Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse  
plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system  
, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

#### Current Chromatogram(s)

