

File : L:\QC Data\GCMS-07 Minimal\DATA\02242201\37 - SAK.D
Operator : Organic QC department
Acquired : 25 Feb 2022 00:31 using AcqMethod KERLA-2017_S100.M
Instrument : 5977B MSD
Sample Name: P-037S-10X
Misc Info :
Vial Number: 90



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

