



=====

Acq. Operator : SYSTEM	Seq. Line :
Sample Operator : SYSTEM	
Acq. Instrument : LCMS-2	Location : P1-B-09
Injection Date : 6/23/2023 5:17:52 PM	Inj : 1
	Inj Volume : 1.000 µl
Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\062323\062323101 2023-06-23 13-48-58\Purity_195-315.M	
Last changed : 2/6/2023 2:02:18 PM by SYSTEM	
Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M	
Last changed : 2/6/2023 2:02:18 PM by SYSTEM	
Method Info : OPERATOR: Organic QC lab	
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.	

Sample Info : Roteneone in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

