

=====

Acq. Operator	: SYSTEM	Seq. Line	: 9
Acq. Instrument	: LCMS-1	Location	: P1-D3
Injection Date	: 5/25/2021 9:18:46 AM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 5/21/2021 4:03:06 PM by SYSTEM  
Analysis Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 5/25/2021 9:11:15 AM by SYSTEM  
(modified after loading)

Method Info : Operator: Oraganic QC lab  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %  
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml  
/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.  
Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV  
Reference.



Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

