

File :D:\MassHunter\GCMS\1\DATA\03192001\14- D  
Operator : Organic QC department  
Acquired : 19 Mar 20 10:00 am using AcqMethod KERLA-2017\_Short\_15Min.M  
Instrument : 5977B MSD  
Sample Name: P-066S  
Misc Info : Mirex in MeOH  
Vial Number: 70



Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

