

Data File : D:\DATA\12171901\06- .D

Sample Name : P-083S-10X



```
=====
Acq. Operator   : SYSTEM                      Seq. Line :    2
Acq. Instrument : LCMS-1                      Location  :    P1-B5
Injection Date  : 12/17/2019 1:01:01 PM      Inj       :    1
                                                Inj Volume: 1.000 µl
Different Inj Volume from Sample Entry! Actual Inj Volume : 0.100 µl

```

```
Method          : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 10/18/2019 4:18:14 PM by SYSTEM
Method Info     : Operator: GA
                  Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                  Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                  /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
                  Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV
                  Reference.
```

Sample Info : Carbaryl in MeOH

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

